



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/651,871	08/29/2003	Horace Winston Hale	HORA.P0101US	9347
23908 7590 08/18/2009 RENNER OTTO BOISSELLE & SKLAR, LLP 1621 EUCLID AVENUE NINETEENTH FLOOR CLEVELAND, OH 44115				
EXAMINER COMSTOCK, DAVID C				
ART UNIT		PAPER NUMBER		
3733				
MAIL DATE		DELIVERY MODE		
08/18/2009		PAPER		

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/651,871

Applicant(s)

HALE ET AL.

Examiner

DAVID COMSTOCK

Art Unit

3733

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 06 May 2009.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-20,41-61 and 63 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-20,41-61 and 63 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 09 June 2006 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☐ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO/SB08)
Paper No(s)/Mail Date _____
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date _____
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: _____

DETAILED ACTION

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(a) the invention was known or used by others in this country, or patented or described in a printed publication in this or a foreign country, before the invention thereof by the applicant for a patent.

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 1-2, 4-11, 14-20, 41-61 and 63 are rejected under 35 U.S.C. 102(a) and 102(e) as being anticipated by Reiley (2002/0123806; cited by Applicant).

Reiley discloses the claimed invention including a facet prosthesis having superior and inferior implants, e.g., 28, 50 and a translaminar fixation system comprising, e.g., screws 32, spikes, surface treatments or other means known in the art (see, e.g., Figs 5 and 7 and paragraphs 0045, 0052, 0058 and 0061). The implant comprises articulating surfaces that are concave and convex and are smooth and configured to articulate in multiple directions while in contact with each other (see, e.g., para. 0053). The device may be formed of common implant materials such as titanium chrome cobalt, polyethylene, ceramic, etc. (see, e.g., para. 0043). Moreover, the device may have a porous coating, biologic treatment, or surface treatment, etc., (see, e.g., para. 0061).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 3, 12, and 13 are rejected under 35 U.S.C. 103(a) as being unpatentable over Reilly (2002/0123806; cited by Applicant).

Reilly discloses the claimed invention except for the explicitly reciting that the inferior implant is configured to interact with the translaminar fixation mechanism such that the translaminar fixation mechanism is offset from about 0 degrees to about 15 degrees and the superior and inferior implants range from about 2 mm thick to about 15 mm thick (although it is noted that the figures would indicate that the portions of the device are in fact approximately within these ranges). Nevertheless, it would have been obvious to one having ordinary skill in the art at the time the invention was made to construct the device of Reilly with the inferior implant being configured to interact with the translaminar fixation mechanism such that the translaminar fixation mechanism is offset from about 0 degrees to about 15 degrees and the superior and inferior implants range from about 2 mm thick to about 15 mm thick, e.g., to accommodate the varying needs of individual patients (note para. 0051 in this regard), since it has been held that where the general conditions of a claim are disclosed in the prior art, discovering the optimum or workable ranges involves only routine skill in the art. In re Aller, 105 USPQ 233.

Response to Arguments

Applicant's arguments pertaining to the claims as amended in the reply filed 06 May 2009 have been considered and are persuasive for the reasons advanced by Applicant, but are moot in view of the newly discovered grounds of rejection, as set forth above.

Conclusion

Applicant's amendment directed to the entire area of the surfaces being smooth, as set forth in the claims, necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to David Comstock whose telephone number is (571) 272-4710 (a detailed message should be left if Examiner is unavailable). If attempts to

reach the Examiner by telephone or voicemail are unsuccessful, the examiner's supervisor, Eduardo Robert, can be reached at (571) 272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

/David Comstock/

Examiner, Art Unit 3733

/Eduardo C. Robert/

Supervisory Patent Examiner, Art Unit 3733